

EC DECLARATION OF CONFORMITY FOR MACHINERY

We, ITE n.v. - Europark Oost 24 - 9100 Sint-Niklaas - Belgium,
hereby declare that the ITE Digital Vacuum Gauge **ITE-55V** is
manufactured in accordance with the following harmonized standards and
directives:

Normative standard EN 61326-1: 2013	SPECIFICATIONS TO WHICH CONFORMITY IS DECLARED	LEVEL
EN 61000-4-2: 2009 Electrostatic Discharge Test	2,4,6,abd 8 kV Air Discharge 2,4, and 6 kV Contact Discharge	A
EN 61000-4-3: 2006 AMD1:2007 AMD2:2010 Radiated Susceptibility Test	4.5 V/M Minimum (4.5 V/M test 900 MHz Cell Phone, 200 Hz, 100% AM	A
EN 55011/EN55022 Radiated Emission	CISPR 16-2-3	B
Low Voltage Directive	2014/35/EU	
EMC Directive	2014/30/EU	



Dirk Maes
Technical Director