

Certificate of Conformity

It is hereby Certified that Advanced Test Products Europe, model number ATPZX-1A, refrigerant leak detector, has been tested to conform to European Norm NEN-EN 14624 - HFC-134a Refrigerant Electronic Leak Detectors, Minimum Performance Criteria.

Testing has been conducted using test equipment as defined by the relevant standards, using an NIST traceable leak source; and the units have met sensitivity, accuracy, and performance requirements defined by the standards.

ATPZX-1A:

Stationary Test Threshold: 1g/yr

Moving Test Threshold: 1g/yr

Date: 05/03/2012

Location: Owatonna, MN, USA



Conducted By: Matt Thompson